

## ***Vibrio cholerae* TaqMan Lyophilized Probe/Primer and Control Set - 100 Reactions**

**Product # TM38510L**

Component	Product # TM38510L (100 Reactions)	Volume upon Reconstitution
<i>Vibrio cholerae</i> Primer & Probe Mix (Lyo)	1 (Lyophilized)	1 X 280 µL
<i>Vibrio cholerae</i> Positive Control (Lyo)	1 (Lyophilized)	1 X 150 µL
Nuclease-Free Water	1 x 1.25 mL	N/A

### **Applications**

Lyophilized TaqMan Probe/Primer and Control Set for *Vibrio cholerae* detection using real-time PCR based on the use of TaqMan® technology.

### **Storage Conditions and Product Stability**

- All kit components should be stored at -20°C upon arrival.
- Once reconstituted, repeated thawing and freezing (>2 times) of the Positive Control should be avoided, as this may affect the performance of the assay. If the reagents are to be used only intermittently, they should be frozen in aliquots.
- All kit components can be stored for 2 years after the date of production without showing any reduction in performance.

### **Precautions and Disclaimers**

- Do not store the kit at room temperature. Store the kit at -20°C upon arrival. Please refer to **Storage Conditions and Product Stability** for further information.
- This product is designed for research purposes only. It is not intended for human or diagnostic use.

### **Customer-Supplied Reagents and Equipment**

- Appropriate Real-Time PCR instrument with FAM and HEX filter channel.

### **Procedure**

- **Reconstitution of *Vibrio cholerae* Primer & Probe Mix (Lyo)**

Note: **(Failure to dissolve the Primer & Probe Mix completely might affect the efficiency of the PCR assay).**

1. Add **275 µL** Nuclease Free Water (provided in the kit) and let it rehydrate at room temperature for 5 minutes.
2. Briefly vortex to dissolve the lyophilized Primer & Probe Mix.
3. Once completely dissolved, briefly spin down the Primer & Probe Mix tube
4. The Primer & Probe Mix is now ready to setup the PCR. Store the reconstituted Primer & Probe Mix at -20°C if not used for PCR setup immediately.

- **Reconstitution of *Vibrio cholerae* Positive Control (Lyo)**

Note: **(Failure to dissolve the Positive Control completely might affect the efficiency of the PCR assay).**

1. Add **120 µL** Nuclease Free Water (provided in the kit) and let it rehydrate at room temperature for 5 minutes.
2. Use a pipette with a sterile tip to dissolve the lyophilized Positive Control.
3. Once completely dissolved, briefly vortex and spin down the Positive Control tube
4. The Positive Control is now ready to setup the PCR. Store the reconstituted Positive Control at -20°C if not used for PCR setup immediately.

- **Recommended Reaction Conditions:**

1. *Vibrio cholerae* Primer Mix (2 µL per 20 µL PCR reaction)
2. *Vibrio cholerae* Positive Control (2-8 µL per 20 µL PCR reaction)

## Description

The positive control is a synthetic oligonucleotide containing a partial *Vibrio cholerae* gene fragment.

**Table 1. Recommended TaqMan PCR Assay**

PCR Cycle	Step	Temperature	Duration
<b>Cycle 1</b>	Step 1	95°C	3 min
<b>Cycle 2 (40x)</b>	Step 1	95°C	15 sec
	Step 2	60°C	30 sec

**Table 2. Interpretation of Assay Results**

FAM (Target detection)	HEX (PCR validation)	Result
+	+	Positive
-	+	Negative
-	-	PCR inhibited

Related Products	Product #
<i>Vibrio cholerae</i> TaqMan PCR Lyophilized Kit	TM38550L
Stool DNA Isolation Kit	27600

## Technical Support

Contact our Technical Support Team between the hours of 9:00 and 5:30 (Eastern Standard Time) at (905) 227-8848 or Toll Free at 1-866-667-4362. Technical support can also be obtained from our website or through email at [techsupport@norgenbiotek.com](mailto:techsupport@norgenbiotek.com).