

Soil RNA Purification Kit - Supplementary Protocol for Isolation of RNA from Waste Water Product # 27750

Customer-Supplied Reagents

- Centricon® Plus-70 Centrifugal Filter with a cut-off of 10 kDa (Merck Millipore)
- Benchtop microcentrifuge
- RNase-free microcentrifuge tubes
- 96-100% ethanol
- 70 % ethanol (Freshly prepared)

Notes Prior to Use

- All centrifugation steps are performed at room temperature.
- A variable speed centrifuge should be used for maximum kit performance. If a variable speed centrifuge is not available a fixed speed centrifuge can be used, however reduced yields may be observed.
- Ensure that all solutions are at room temperature prior to use.
- Prepare a working concentration of **Wash Solution A** by adding 42 mL of 96 - 100 % ethanol (provided by the user) to each supplied bottle containing the concentrated **Wash Solution A**. This will give a final volume of 60 mL. The labels on the bottles have a box that may be checked to indicate that the ethanol has been added.
- This kit is provided with 2 separate columns. When columns are removed from the labelled bags they are supplied in they can easily be identified as follows:
 - Humic Acid Removal Columns – column has blue and white contents
 - Spin Columns – column has grey and white contents

1. Sample Treatment

- a) Concentrate the sample using the Centricon® Plus-70 Centrifugal Filter with a cut-off of 10 kDa (Merck Millipore).
- b) Transfer the concentrated sample (approximately 250 µL) from the concentrate collection cup into an RNase-free microcentrifuge tube (provided by the user).

2. Lysate Preparation

- a) Add 50 µL of **Solution BX** and add 50 µL **Binding Buffer E** sequentially. Mix by inverting the tube a few times.
- b) Incubate for 5 minutes on ice.
- c) Spin the lysate for 1 minute at **20,000 × g (~14,000 RPM)** to pellet any protein and soil particles. If the lysate has a clear colour, skip Step 2d.
- d) Using a pipette, transfer up to 450 µL of supernatant into a **Humic Acid Removal Column (blue and white contents)** without any contact with the pellet.
- e) Spin the column at **4,000 × g (~8,000 rpm)** for 30 seconds. **Don't discard the flow through that contains RNA.**
- f) Using a pipette, transfer up to 400 µL of supernatant (avoid any contact with the pellet when collecting the supernatant) into a RNase-free microcentrifuge tube (not provided).
- g) Add 300 µL of **Lysis Buffer QP** and 700 µL of **70 % ethanol** (provided by the user) to the lysate collected above. Vortex to mix. **Proceed to Step 3.**

3. Binding to Column

- a) Assemble a **Spin Column (grey and white contents)** with one of the provided collection tubes.

- b) Apply up to 650 μ L of the clarified lysate with ethanol onto the column and centrifuge for 1 minute at **20,000 \times g (~14,000 RPM)**. Discard the flowthrough and reassemble the spin column with the collection tube.

Note: Ensure the entire lysate volume has passed through into the collection tube by inspecting the column. If the entire lysate volume has not passed, spin for an additional minute at **20,000 \times g (~14,000 RPM)**.

- c) Repeat step 3b with the remaining lysate.

4. Column Wash

- a) Apply 500 μ L of **Binding Buffer B** to the column and centrifuge for 1 minute at **20,000 \times g (~14,000 RPM)**.
- b) Discard the flowthrough and reassemble the spin column with its collection tube.
- c) Apply 500 μ L of **Wash Solution A** to the column and centrifuge for 1 minute at **20,000 \times g (~14,000 RPM)**. Discard the flowthrough and reassemble the spin column with its collection tube.
- d) Repeat **4c**.
- e) Spin the column for 2 minutes at **20,000 \times g (~14,000 RPM)** in order to thoroughly dry the resin. Discard the collection tube.

5. RNA Elution

- a) Place the column into a fresh 1.7 mL Elution tube provided with the kit.
- b) Add 50 μ L of **Elution Solution A** to the column.
- c) Centrifuge for 2 minutes at **200 \times g (~1,500 RPM)**, followed by a 1 minute spin at **20,000 \times g (~14,000 RPM)**. Note the volume eluted from the column. If the entire volume has not been eluted, spin the column at **20,000 \times g (~14,000 RPM)** for 1 additional minute.

6. Storage of RNA

The purified RNA may be stored at -20°C for a few days. It is recommended that samples be placed at -70°C for long term storage.

Technical Support

Contact our Technical Support Team between the hours of 8:30 and 5:30 (Eastern Standard Time) at (905) 227-8848 or Toll Free at 1-866-667-4362.

Technical support can also be obtained from our website (www.norgenbiotek.com) or through email at techsupport@norgenbiotek.com.

3430 Schmon Parkway, Thorold, ON Canada L2V 4Y6
Phone: (905) 227-8848
Fax: (905) 227-1061
Toll Free in North America: 1-866-667-4362